

Moving  
Rehabilitation  
Forward™



# TRITON®

## User Manual

Model 4749 - Triton® Traction Unit  
Model 4798 - Triton® DTS Traction Unit

 **chattanooga™**  
ISO 13485 Certified

**Electromagnetic  
Compatibility  
(EMC) Tables**

**TOC**





# TABLE OF CONTENTS

- FOREWORD ..... 1**
  - ▶ Product Description ..... 1
- ABOUT TRACTION THERAPY ..... 2-8**
  - ▶ Precautionary Instructions ..... 2
  - ▶ Cautions ..... 3-4
  - ▶ Warnings ..... 4-5
  - ▶ Dangers ..... 5
  - ▶ Overview of Traction Therapy ..... 6
  - ▶ Common Terms ..... 6
  - ▶ Indications ..... 7
  - ▶ Contraindications ..... 7
- NOMENCLATURE ..... 8-13**
  - ▶ Base Unit - Front View ..... 8-9
  - ▶ Base Unit - Rear View ..... 10-11
  - ▶ LCD Base ..... 12
  - ▶ Symbol Definitions ..... 13
- SPECIFICATIONS ..... 14-16**
  - ▶ Traction Unit Dimensions and Specifications ..... 14
  - ▶ sEMG Module Dimensions and Specifications ..... 15
  - ▶ Description of Device Markings ..... 16
- SETUP ..... 17-23**
  - ▶ Contents of Carton ..... 17
  - ▶ Mounting Unit on Table ..... 18
  - ▶ Attaching the Patient Switch Hook ..... 19
  - ▶ Installing the sEMG Module ..... 20
  - ▶ Removing the sEMG Module ..... 21
  - ▶ Initial Unit Setup ..... 22-23
- OPERATION ..... 24-48**
  - ▶ Patient Preparation ..... 24-25

- ▶ Operating Controls - Home Screen ..... 26-28
- ▶ Operating Controls - Treatment Screen ..... 29-32
- ▶ sEMG Traction Treatment ..... 33-35
- ▶ Starting, Pausing and Stopping Treatment ..... 36-37
- ▶ Saving Treatment to a Patient Data Card ..... 38-40
- ▶ Retrieving Data from Existing Patient Data Card ..... 41
- ▶ Erasing Patient Data Card ..... 42
- ▶ User Protocols ..... 43-44
- ▶ Clinical Protocols™ ..... 45-46
- ▶ Multimedia Card (MMC) Graphics Library ..... 47
- ▶ DTS Pull Patterns (If Applicable) ..... 48
- ACCESSORIES ..... 49-50**
- TROUBLESHOOTING ..... 51-54**
  - ▶ Troubleshooting ..... 51-52
  - ▶ Error Messages ..... 53-54
- MAINTENANCE ..... 55-56**
  - ▶ Cleaning ..... 55
  - ▶ Preventative Maintenance Schedule ..... 55
  - ▶ Calibration Requirements ..... 55
  - ▶ Service ..... 56
- WARRANTY ..... 57**





# FOREWORD

This manual has been written for the operators of the Triton traction unit. It contains general instructions for operation, precautionary instructions, and maintenance recommendations. In order to obtain maximum life and efficiency from your Triton traction unit, and to assist in the proper operation of the unit, read and understand this manual thoroughly.

The specifications put forth in this manual were in effect at the time of publication. However, owing to DJO, LLC's policy of continuous improvement, changes to these specifications may be made at any time without obligation on the part of DJO, LLC.

Before administering any treatment to a patient, you should become acquainted with the operating procedures, as well as the indications, contraindications, cautions, warnings, and dangers. Consult other resources for additional information regarding the application of traction therapy.

## Product Description

The Triton traction unit is a simple to use digital touch screen user interface traction unit that offers static, intermittent, and cyclic traction with user definable hold, rest, and treatment times. An optional sEMG Module and DTS software upgrade are available for separate purchase. Patient Pain Profiles are recorded before and after treatment with a Numeric Pain Scale, VAS (Visual Analog Scale), and Pain Map for each patient. Once treatment parameters are customized for a patient, clinicians can store them as either the default user protocol or as one of twenty user defined protocols.

**This equipment is to be used only under the prescription and supervision of a licensed practitioner.**





# ABOUT TRACTION THERAPY

## PRECAUTIONARY INSTRUCTIONS

The precautionary instructions found in this section and throughout this manual are indicated by specific symbols. Understand these symbols and their definitions before operating this equipment. The definitions of these symbols are as follows:



### Caution-

Text with a “CAUTION” indicator will explain possible safety infractions that could have the potential to cause minor to moderate injury or damage to equipment.



### Warning-

Text with a “WARNING” indicator will explain possible safety infractions that will potentially cause serious injury and equipment damage.



### Danger-

Text with a “DANGER” indicator will explain possible safety infractions that are imminently hazardous situations that would result in death or serious injury.

### NOTE:

Throughout this manual “NOTE” may be found. These Notes are helpful information to aid in the particular area or function being described.



### Explosion Hazard

Text with an “Explosion Hazard” indicator will explain possible safety infractions if this equipment is used in the presence of flammable anesthetics.



## ABOUT TRACTION THERAPY



### CAUTION

- Read, understand, and practice the precautionary and operating instructions. Know the limitations and hazards associated with using sEMG, DTS Pull Patterns and the Triton traction unit. Observe the precautionary and operational decals placed on the unit.
- DO NOT operate this unit in an environment where other devices are being used that intentionally radiates electromagnetic energy in an unshielded manner. Portable and mobile RF communications equipment can affect Medical Electrical Equipment.
- This equipment generates, uses, and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to other devices in the vicinity. However, there is no guarantee that interference will not occur in a particular installation. Harmful interference to other devices can be determined by turning this equipment on and off. Try to correct the interference using one or more of the following: reorient or relocate the receiving device, increase the separation between the equipment, connect the equipment to an outlet on a different circuit from that which the other device(s) are connected and consult the DJO, LLC Service Department for help.
- This unit should be operated, transported, and stored in temperatures between -18 °C and 60 °C (0° F and 140° F).
- The unit should be routinely checked before each use to determine all controls function normally.
- Do not use electrodes that have been previously used as the electrodes used with this unit are designed for single use only.



### CAUTION

- Do not operate the unit when connected to any unit other than Chattanooga devices. Do not use devices manufactured by other companies on Chattanooga equipment. DJO, LLC is not responsible for any consequence resulting from using products manufactured by other companies.
- Handle the unit with care. Inappropriate handling of the unit may adversely affect its characteristics.
- Before each use, inspect the Traction Cord for wear. Prolonged wear on the cord will cause it to break, which may cause sudden release of traction pressure on a patient.
- Inspect lead wires and associated connectors for signs of damage before each use. Replace damaged lead wires immediately before any treatment is applied.
- Always test the Patient Interrupt Switch cable before each use for proper operation.
- Do not use sharp objects such as a pencil point or ballpoint pen to operate the buttons on the Touch Screen base as damage may result.
- Do not remove the cover. This may cause unit damage, malfunction, electrical shock, fire, or personal injury. There are no user-serviceable parts inside the unit. If a malfunction occurs, discontinue use immediately, disconnect the Mains Power Cord from the outlet, and consult the dealer for repair service.
- Do not disassemble, modify, or remodel the unit or accessories. This may cause unit damage, malfunction, electrical shock, fire, or personal injury.
- Do not use the traction unit near devices such as X-ray units or diathermy units. These units may emit high frequency noise that may affect the operation of the unit.



## ABOUT TRACTION THERAPY



### CAUTION

- Do not use the Clevis as a handle to pick up or carry the unit.
- Do not permit any foreign materials or liquids to enter the unit. Take care to prevent any foreign materials including, but not limited to, inflammables, water, and metallic objects from entering the unit. These may cause unit damage, malfunction, electrical shock, fire, or personal injury.
- If you have difficulty operating the unit after carefully reviewing this operator's guide, contact your DJO, LLC dealer for assistance.
- Failure to use and maintain the traction unit and its accessories in accordance with the instructions outlined in this manual will invalidate your warranty.
- The Patient Interrupt Switch is essential for the functioning of the unit. If it is not connected, or if it is malfunctioning, the unit will not work.
- A licensed practitioner experienced with traction therapy must be familiar with all instructions contained in this manual before administering traction therapy. Do not attempt to become familiar with the Triton traction unit while administering traction therapy on a patient.
- In the event of a loss of power to the unit or when quick release is needed, traction tension should only be released by having the patient move towards the traction head to release the tension on the rope. Once the tension on the rope has been released, loosen the patient harness adjustment straps.



### WARNING

- This device should only be used under the continued supervision of a licensed practitioner.
- Care must be taken when operating this unit adjacent to or stacked with other equipment. Potential electromagnetic or other interference could occur to this or other equipment. Try to minimize this interference by not using other equipment (i.e. cell phones, etc.) in conjunction with it.
- Use of controls or adjustments or performance of procedures other than those specified herein may result in a hazardous traction related injury.
- Before connecting the unit to an electrical outlet, make certain the unit is electronically grounded by connecting only to a grounded electrical service receptacle conforming to the applicable national and local electrical codes.
- Do not use a damaged Mains Power Cord. Using a damaged Mains Power Cord may cause unit damage, malfunction, electrical shock, fire, or personal injury. If the Mains Power Cord becomes damaged, discontinue use immediately and contact the dealer for replacement of the Mains Power Cord.
- To prevent electrical shock, disconnect the unit from the power source before attempting any maintenance procedures.
- This device should be kept out of the reach of children.
- Always hand tighten the Lock Knob securely to avoid any slippage.
- The Patient Interrupt Switch must be in the patient's grasp throughout the traction therapy.
- Do not apply external pressure (heat or cold) to the electrode sites during therapy.
- Dispose of all products in accordance with local and national regulations and codes.





# ABOUT TRACTION THERAPY



## WARNING

- Use only tables, stands, power cords and accessories that are specially designed for the Triton traction unit. Do not use accessories manufactured by other companies on the Triton traction unit. DJO, LLC is not responsible for any consequence resulting from using products manufactured by other companies. The use of other accessories or cables may result in increased emissions or decreased immunity of the Triton traction unit.
- Do not use other accessories other than those supplied with the system or recommended by DJO, LLC. The safety of other products has not been established, and their use could result in injury to the patient.
- To prevent accidental disengagement, this unit must be securely attached to the mounting surface of the pedestal or traction stand. It is the responsibility of the user to verify the adequacy of the installation before use in therapy.
- Do not apply electrodes over broken or compromised skin (e.g., sunburn, cuts, acne) due to increased risk of skin reactions.
- Do not plug lead wires into power outlets such as wall sockets and line cord receptacles. Doing so could result in severe shock or burns whether or not the lead wires are attached to the unit.
- In the event that an Error message or Warning appears beginning with a 2 or 3, immediately stop all use of the system and contact the dealer or DJO, LLC for service. Errors and Warnings in these categories indicate an internal problem with the unit that must be tested by DJO, LLC or a Field Service Technician certified by DJO, LLC before any further operation or use of the unit. Use of a unit that indicates an Error or Warning in these categories may pose a risk of injury to the patient, user or extensive internal damage to the unit.
- Disconnect the traction unit from the power source before attempting any maintenance, installation, removal, or replacement procedures to prevent electrical shock and possible damage to the unit.
- The traction unit should only be used by a qualified licensed practitioner in a position of supervision during therapy. For that reason, do not attempt to put yourself in traction with this unit.



## DANGER

- Handle, clean, and dispose of components and accessories that have come in contact with bodily fluids according to National, Local and Facility rules, regulations and procedures.
- Do not connect the unit to an electrical supply without first verifying that the power supply is the correct voltage. Incorrect voltage may cause unit damage, malfunction, electrical shock, fire, or personal injury. Your unit was constructed to operate only on the electrical voltage specified on the Voltage Rating and Serial Number plate. Contact your DJO, LLC dealer if the unit is not properly rated.
- Should Traction Cord fray or damage to the Traction Cord or knot be apparent from visual inspection, immediately stop use of the unit and contact the Dealer or DJO, LLC for service.
- Do not attempt to repair the Traction Cord.
- An improperly tied knot may result in injury. Do not attempt to re-tie the knot unless properly trained.
- Not for use in the presence of flammable anesthetics.





# ABOUT TRACTION THERAPY

## OVERVIEW OF TRACTION THERAPY

### Effects of Traction Therapy

The Triton traction device provides a treatment in static, intermittent, and cyclic distraction forces to relieve pressures on structures that may be causing pain of skeletal or muscular origin (cervical, thoracic, lumbar, hip, wrist, shoulder). Therapeutic distraction can be applied in a variety of programmable patterns, cycles and functions.

### COMMON TERMS

#### Progressive Traction

Progressive traction refers to a traction phase during the treatment when the tension gradually increases.

#### Regressive Traction

Regressive traction refers to a traction phase during the treatment when the tension gradually decreases.

#### TX (Traction)

TX (Traction) refers to the type of traction used during a treatment. There are three modes to choose from: Static, Intermittent or Cyclic Traction Mode.

#### Steps

The term “step” refers to the increment in which the traction tension is either increased or decreased during therapy.

### Static Traction Therapy

This term denotes that a steady amount of traction is applied for periods from a few minutes up to 99 minutes. The shorter duration is usually coupled with more tension. Static lumbar traction is most effective if a split table is utilized to reduce friction. It is important that it is the type that maintains constant tension. This way, any slack developed as the patient relaxes during the traction therapy is automatically taken up and the desired amount of traction is maintained. Static traction is sometimes referred to as sustained traction.

### Intermittent Traction Therapy

This form of traction alternates traction tension between tension levels - Maximum and Minimum - every few seconds throughout the timed treatment. It is also most effective if a split table is used to reduce friction when giving lumbar traction. In the progressive and regressive phases, the traction unit pulls to the calculated tension, holds for the set hold time, then drops to 50% of the tension level, holds the set rest time and then repeats this step for the number of steps selected. However, once the minimum level is reached, the traction unit uses the minimum level for the rest time.

### Cyclic Traction

Cyclic traction refers to progressive and regressive phases of the traction program being repeated continuously throughout the entire course of the traction treatment.







# ABOUT TRACTION THERAPY

## INDICATIONS

The Triton traction device provides traction and mobilization of skeletal structures and skeletal muscles.

The Triton traction device may be used to relieve peripheral radiation/ sciatica and pain associated with:

- Protruding discs
- Bulging discs
- Herniated discs
- Degenerative disc disease
- Posterior facet syndrome
- Acute facet problems
- Radicular pain
- Prolapsed discs
- Spinal root impingement
- Hypomobility
- Degenerative joint disease
- Facet syndrome
- Compressions fracture
- Joint pain
- Discogenic pain

The Triton traction device achieves these effects through mechanical traction of intervertebral discs, that is, unloading due to distraction and positioning.

## CONTRAINDICATIONS

**Traction is contraindicated for the following:**

- Structural disease secondary to tumor or infection
- Patients with vascular compromise
- Any condition for which movement is contraindicated
- Patients with acute strains, sprains, and inflammation which would be aggravated by traction therapy
- Patients with joint instability of the spine
- Pregnancy
- Osteoporosis
- Hiatus hernia
- Claustrophobia
- Cardiac or pulmonary problems





## BASE UNIT- FRONT VIEW


Touch Screen  
User Interface

Traction Cord

Clevis  
Attachment Slots

sEMG  
Lead Wire  
Receptacle  
(Model 4798  
or with sEMG  
Module P/N  
4766)

Patient  
Interrupt  
Switch  
Receptacle



The Base Unit serves to house the mechanical and electrical components that provide the actual traction tension for the Triton traction unit under the electronic control and supervision of the Controller.



### Traction Cord

To prevent wear and fraying, the unit should be mounted facing toward the desired direction of pull so that the Traction Cord does not contact the side of the slot from which it extends.

## DANGER

- Should Traction Cord fray or damage to the Traction Cord or knot be apparent from visual inspection, immediately stop use of the unit and contact the Dealer or DJO, LLC for service.
- Do not attempt to repair the Traction Cord.
- An improperly tied knot may result in injury. Do not attempt to re-tie the knot unless properly trained.



# NOMENCLATURE

## BASE UNIT- FRONT VIEW (CONTINUED)



### Accessory Clip

The Accessory Clip allows you to attach the desired accessory (i.e., Spreader Bar or traction harnesses) to the unit.



### Patient Interrupt Switch Receptacle

The Patient Interrupt Switch Receptacle is located on the front of the unit.



### Clevis Attachment Slots

The Clevis Attachment Slots serve as a connection point for other traction apparatus (not included), such as cervical or wrist.



## CAUTION

- The Patient Interrupt Switch is essential for the functioning of the unit. If it is not connected, or if it is malfunctioning, the unit will not work.



## CAUTION

- Do not use the Clevis as a handle to pick up or carry the unit.




### sEMG Lead Wire Receptacle (Model 4798 or with sEMG Module P/N 4766)

The sEMG lead wire connection is located on the front of the unit.




## BASE UNIT- REAR VIEW



The rear view of the base unit is the site of electric current control and surge protection, and serves as a connection point for both the traction controls and safety features of the Triton traction unit.



**Power On/Off Switch**   
The Power On/Off Switch is a toggle switch located on the back of the unit. This switch controls the flow of electricity from the outlet to the unit.



**Mains Power Cord Receptacle**  
The Mains Power Cord Receptacle accepts the Mains Power Cord (female end).




**Screen Contrast Control**  
The Screen Contrast Control is used to set a comfortable viewing of brightness of the Touch Screen User Interface. **Functional On Monochrome Units Only.**





# NOMENCLATURE

## BASE UNIT- REAR VIEW (CONTINUED)



### **Lock Knob**


The Lock Knob is used to clamp the Triton traction unit to the base (i.e. pedestal, traction stand).



# NOMENCLATURE

## LCD BASE

The LCD Base serves as a programming terminal during the selection of the traction parameters, as well as a display showing all the factors affecting the traction during therapy. Traction parameters are selected with the buttons on the LCD Base and the Touch Screen User Interface. The Touch Screen and beeper make various audio and visual indications, and warn the operator when unsuitable parameters are chosen.



### (1) LED Indicator (Power On/Off)

This indicator will illuminate when the unit is powered on.

**NOTE:** The LED indicator will blink when the unit goes in screen saver mode (the screen will go blank after twenty minutes of inactivity). Simply touch the screen to reactivate.

### (2) Clinical Resources

This button allows you to access the following features of the traction unit:

- **Patient Card**
  - Patient Card (Patient Name)
  - Edit Current Pain Profile
  - Edit/Save Completed Pain Profile
  - Erase Patient Card
- **Protocols**
  - Retrieve Protocol
  - Save Protocol
  - Clinical Protocols™
- **Utilities**
  - Unit Settings

### (3) Stop

This button will stop the treatment program. Traction tension slowly decreases to zero.

### (4) Pause

This button will pause the treatment program. Traction tension slowly decreases to zero. Press the pause button again to resume treatment program and traction tension.

### (5) Start

Touch this button to start the treatment program.

### (6) Patient Data Card Port

Access port to insert a Patient Data Card to save and retrieve patient treatments.

### (7) Multimedia Card (MMC) Port

Access port to insert a MMC Card to display traction techniques and the anatomical library.

TOC





# NOMENCLATURE

## SYMBOL DEFINITIONS

Below are the definitions for all of the Symbols used on the Triton traction unit hardware and software. Study and learn these symbols before any operation of the unit.

### System Hardware Symbols



Contrast Control



On/Off Switch



Data Port (Factory Use Only)



Lock/Unlock Lock Knob



Stop Treatment



Pause Treatment



Start Treatment



sEMG Lead Wires



Patient Card/ Multi-Media Card (MMC)/ DTS Software Upgrade Card



Patient Interrupt Switch



Clinical Resources

### System Software Symbols



Move UP



Move DOWN



Accept and Return



Cancel



Back



Forward



Home

TOC





# SPECIFICATIONS


## TRACTION UNIT DIMENSIONS AND SPECIFICATIONS


**Width** .....24 cm (9.5 in)  
**Depth** .....45 cm (17.5 in)  
**Height** .....45 cm (17.5 in)

### Weight

Standard Weight .....14 kg (30 lbs)  
 Shipping Weight .....18 kg (40 lbs)

### Power

Voltage .....100V-240V (50/60Hz)  
 Duty Cycle ..... Continuous Operation  
 Power Consumption .....75 VA  
 Current Consumption .....3.2 Amps Max  
 Electrical Class ..... Class I  
 Traction Electrical Type .....Type B 

sEMG Electrical Type ..... Type BF 

### Traction Modes

Static, Intermittent, Cyclic, and their combinations.

**Traction Type:** Mechanical

**LCD Display:** High contrast color touch screen

### Traction Tension

Parameters	Minimum	Maximum	Increments
Traction Period	1 minute	99 minutes	1 minute
Hold Period	0 seconds	99 seconds	1 second
Rest Period	0 seconds	99 seconds	1 second
Traction Tension	0 kg (0 lb/0 N)	90 kg (200 lb/890 N)	1 kg (1 lb/5 N)
Progressive and Regressive Steps	1 step	9 steps	1 step



Meets Directive 93 /42 /EEC  
 IEC/UL/EN: 60601-1 60601-1-2



Intertek  
 3187214

Certified to CAN/CSA  
 Standard C22.2 No. 601.1-M90w/A2



TOC





# SPECIFICATIONS

## SEMG MODULE DIMENSIONS AND SPECIFICATIONS

<b>Width</b> .....	16.83 cm (6.63 in)
<b>Depth</b> .....	7.62 cm (3 in)
<b>Height</b> .....	4.13 cm (1.63 in)
<b>Input Impedance</b> .....	> 1,000,000 ohm
<b>Input Sensitivity</b> .....	< 1.0 $\mu$ V RMS
<b>Frequency Range</b> .....	15Hz - 1000Hz with CMMR of > 120 dB CMMR at 50/60 Hz > 180 dB
<b>Weight</b> .....	146 g (5.2 oz)

### Product Type

Internally Powered .....	Triton Traction Unit
Electrical Class .....	Class I
Electrical Type .....	Type BF 
For Continuous Operation	

### Regulatory Compliance




Meets Directive 93 /42 /EEC  
IEC/UL/EN: 60601-1 60601-1-2



Certified to CAN/CSA  
Standard C22.2 No. 601.1-M90w/A2

Intertek  
3187214



# SPECIFICATIONS

## DESCRIPTION OF DEVICE MARKINGS

The markings on the Triton traction unit are your assurance of its conformity to the highest applicable standards of medical equipment safety and electromagnetic compatibility. One or more of the following markings may appear on the device:

 0473 Meets Directive 93 /42 /EEC  
IEC/UL/EN 60601-1, 60601-1-2



***This product complies with WEEE Directive 2002/96/EG (waste electrical and electronic equipment). Separate collection for electrical and electronic equipment.***



Certified to CAN/CSA  
Standard C22.2 No. 601.1-M90w/A2



Refer to Instruction Manual/Booklet



Type B Equipment



Type BF Equipment



Protective Earth



## CONTENTS OF CARTON

Remove the Triton traction unit and all accessories from the shipping cartons. Visually inspect for damage.

Report any damage to the carrier immediately.

### Contents of Cartons:

#### Model 4749 - Triton Traction Unit

- Triton Traction Unit
- Patient Data Cards (5)
- Patient Interrupt Switch - 2.75 m (9 ft) in cable length, shielded, 26 AWG
- Patient Interrupt Switch Hook Kit
- Mains Power Cord - 203 cm (80 in) in cable length, shielded, 18 AWG
- User Manual (CD-ROM)

#### Model 4798 - DTS Traction Unit

- Triton DTS Traction Unit
- Patient Data Cards (5)
- sEMG lead wire - 2.44 m (8 ft) in cable length, unshielded, 24/28 AWG
- (4) Dura-Stick II Electrodes - 3 cm (1.25 in) Round
- (4) Dura-Stick II Electrodes - 5 cm (2 in) Round
- Patient Interrupt Switch - 2.75 m (9 ft) in cable length, shielded, 26 AWG
- Patient Interrupt Switch Hook Kit
- Mains Power Cord - 203 cm (80 in) in cable length, shielded, 18 AWG
- User Manual (CD-ROM)

**NOTE:** When shipping the unit back to the dealer or factory, make certain the original packaging is used. If the original packaging is not available, contact DJO, LLC to obtain the following packaging materials for shipment:

48059	Inner Pack Box Top
48093	Left Inner Pack
48094	Right Inner Pack
48092	Shipping Box
48095	Shipping Bag

Any damage sustained from improper packaging may render the warranty null and void.



# SETUP

## MOUNTING UNIT ON TABLE

The Triton traction unit was designed to be utilized with Chattanooga traction tables:



### CAUTION

- Do not operate the unit when connected to any unit other than Chattanooga devices. Do not use devices manufactured by other companies on Chattanooga equipment. DJO, LLC is not responsible for any consequence resulting from using products manufactured by other companies.



### WARNING

- To prevent accidental disengagement, this unit must be securely attached to the mounting surface of the pedestal or traction stand. It is the responsibility of the user to verify the adequacy of the installation before use in patient therapy.

**NOTE:** The Triton traction unit should be mounted facing the direction of pull so the Traction Cord does not contact the side of the slot from which it extends.



- To secure the traction unit to the mounting surface, first loosen the Clamp Knob by turning to the left. Fit the unit to the mounting surface, then turn the Clamp Knob to the right, tightening the clamp and securing the unit in place.
- Remove the cover on the back of the unit and connect the Mains Power Cord (female end) to the Mains Power Cord Receptacle.
- Verify that there is a good connection between the Mains Power Cord and the Mains Power Cord Receptacle. Always make certain that the Mains Power Cord is properly plugged into the unit.
- Check the voltage rating on the serial plate located on the back of the unit. Plug the Mains Power Cord into a 100V-240V AC outlet, as required. Replace the cover on the back of the unit.

# SETUP

## ATTACHING THE PATIENT INTERRUPT SWITCH HOOK

The Patient Interrupt Switch Hook may be placed on either the left or right side of the unit to hook the Patient Interrupt Switch.



A screw hole cap with the Chattanooga logo is located on both sides of the traction unit.

Using a screwdriver, pop the screw hole cap off.



Place the Patient Interrupt Switch Hook over the screw hole and tighten the Hook Screw in place.



Insert the Patient Interrupt Switch Hook Cover to cover the hook screw hole.




The Patient Interrupt Switch can now be placed on the Patient Interrupt Switch Hook.



# SETUP

## INSTALLING THE sEMG MODULE

**NOTE:** Model 4749 can be upgraded to become a traction unit with sEMG by attaching a sEMG Module (P/N 4766) to the unit.



# WARNING


- Disconnect the traction unit from the power source before attempting any maintenance, installation, removal, or replacement procedures to prevent electrical shock and possible damage to the unit.




Turn the unit off. Disconnect the traction unit from its power source. Disconnect the Patient Interrupt Switch from its receptacle.



Gently press the tabs underneath the unit, pull and remove the front panel from the traction unit



Gently push the sEMG Module into the front of the traction unit. Once the plastic of the unit and the module are aligned, the connector should be inserted properly, with the tabs locked in place.



Plug the traction unit back into the power source. Connect the Patient Interrupt Switch and the sEMG lead wire into the receptacles. Turn the unit on.

The traction unit should automatically recognize the sEMG Module with the sEMG Traction button now being active on the Home Screen.

TOC





## REMOVING OF THE sEMG MODULE

### **WARNING**

- Disconnect the traction unit from the power source before attempting any maintenance, installation, removal, or replacement procedures to prevent electrical shock and possible damage to the unit.



Turn the unit off. Disconnect the traction unit from its power source. Disconnect the Patient Interrupt Switch and the sEMG lead wire from the receptacles.




Gently press the tabs underneath the unit, pull and remove the sEMG Module from the traction unit



Gently push the front panel back into the front of the traction unit. Once the plastic of the unit and the front panel are aligned, the connector should be inserted properly, with the tabs locked in place.

Plug the traction unit back into the power source. Connect the Patient Interrupt Switch into its receptacle. Turn the unit on.




The traction unit should automatically recognize that the sEMG Module has been removed, with the sEMG Traction button now being inactive on the Home Screen.






## INITIAL UNIT SETUP

### Accessing User Utilities



Turn the unit On.

Press the Clinical Resources button.



Press the Unit Settings button under the Utilities section.

### Clinic Name

Press the Clinic Name button. Enter the Clinic Name on the keyboard. To save the Clinic Name as entered, press the Return Arrow button. To discard the Clinic Name, press the Cancel (X) button.

**NOTE:** The default of the Clinic Name is set as *Chattanooga*.

### Volume

Press the Volume button until the desired unit volume is achieved. There are six settings: Off, X-Low, Low, Med, High and X-High. Each time the Volume button is pressed, the setting displayed will also give three audible tones of that level.

**NOTE:** The default of the unit volume is set at Medium.

### Units of Measure

There are three units of measure to choose from: kilograms (kg), pounds (lbs), and newtons (N). Press the Unit of Measure button to toggle to the desired unit of measure and the setting will automatically save.

**NOTE:** The default of the unit of measure is set in kilograms (kg).

### Set Date and Time

Press the Set Date and Time button. Press the Move Up or Down Arrow button for the respective area (Year, Month, Day, Hour and Minute) until the desired change is displayed.

**NOTE:** The time is set in a 24-hour format.







## INITIAL UNIT SETUP (CONTINUED)

### Language

To change the language displayed on the unit, press the Language button until the desired language is displayed and the setting will automatically save.

The available languages are:

- English
- German
- French
- Italian
- Spanish
- Dutch
- Portuguese
- Danish
- Norwegian
- Swedish
- Finnish
- Turkish
- Greek
- Chinese
- Czech
- Polish

**NOTE:** The default of Language is set as English.

### Reset Unit Settings

Press the Reset Unit Settings button to restore the unit defaults (Clinic Name, Volume, Unit of Measure, and Language). This control will not change the Date and Time nor will it affect any of the User Protocols stored in the unit.

### Reset Default Protocols

Press the Reset Default Protocols to restore all protocols to factory settings.

**NOTE: All User Protocols will be removed.**

To return to the unit Home Screen, press the Home button.





# OPERATION

## PATIENT PREPARATION

Traction therapy may be given by or on order of a qualified licensed practitioner. The person administering the treatment must be familiar with the principles of traction therapy and be able to choose the correct mode of traction, traction forces, and treatment times. Any settings or types of traction mentioned in this manual are for illustrative and expository purposes only. Each patient should be individually assessed by a licensed practitioner to determine the appropriateness of the parameter settings prior to use.



### CAUTION

- A licensed practitioner experienced with traction therapy must be familiar with all instructions contained in this manual before traction therapy. Do not attempt to become familiar with the Triton traction unit while administering traction therapy on a patient.

To prepare the patient for traction therapy, do the following:



1. Attach the Patient Interrupt Switch by inserting the male end of the cable into the Patient Interrupt Switch Receptacle located on the front of the unit.

**NOTE:** The Patient Interrupt Switch must be plugged in for the unit to operate. If the Patient Interrupt Switch is not inserted into the traction unit, a continuous audible tone will sound and a message will display saying, "**Patient Switch is not plugged into the unit. Correct the problem.**" Touch the Screen anywhere to continue.

2. Turn the unit on from the Power On/Off Switch, located on the back of the unit. Wait about five seconds for the unit to go through initialization before attempting any further patient preparation procedures.
3. Test the Patient Interrupt Switch by pressing the red button. A warning message will be displayed saying *The Patient Interrupt Switch has been pressed. Touch the screen to continue.*



### CAUTION

- Always test the Patient Interrupt Switch cable before each use for proper operation.

4. Give the Patient Interrupt Switch to the patient, instructing that pressing the red button will stop the treatment.



### WARNING

- The Patient Interrupt Switch must be in the patient's grasp throughout the traction therapy.



# OPERATION

## PATIENT PREPARATION (CONTINUED)

5. Position the patient on the appropriate type of table in accordance with the instructions supplied with the table.



### WARNING

- Use only tables, stands, power cords and accessories that are specially designed for the Triton traction unit. Do not use accessories manufactured by other companies on the Triton traction unit. DJO, LLC is not responsible for any consequence resulting from using products manufactured by other companies. The use of other accessories or cables may result in increased emissions or decreased immunity of the Triton traction unit.

6. Fit the appropriate traction harness on the patient by following the fitting instructions supplied with the type of harness to be used.
7. Tighten any slack in the harness that may have occurred during positioning so that it fits the patient according to the manufacturer's instructions.
8. Push and hold the Rope Release button on the Touch Screen and slowly pull the end of the Traction Cord out from the traction unit.
9. Attach the Accessory Clip to the connection point of the traction harness, according to the manufacturer's instructions.




### CAUTION

- In the event of a loss of power to the unit or when quick release is needed, traction tension should only be released by having the patient move towards the traction head to release the tension on the rope. Once the tension on the rope has been released, loosen the patient harness adjustment straps.



# OPERATION

## OPERATING CONTROLS - HOME SCREEN



### Cervical Traction

The Cervical Traction button will go straight to the Cervical Traction menu. The Cervical Traction menu offers the following options:

**Traction Rationale** - Button provides a description of the traction being used as well as definitions of terms used in areas of the Treatment Screen.

**Traction Technique** - Button provides visual steps as to how to set up a patient for cervical traction.

**NOTE:** After viewing the steps for set up, the Forward Arrow button will automatically go to the Treatment Screen.

**Edit** - Button provides access to the Treatment Screen to set up parameters for treatment.

**Back and Home** - Buttons go back to the Home Screen.

**Rope Release** - Releases the Traction Cord from the unit.

### Lumbar Traction

The Lumbar Traction button will go straight to the Lumbar Traction menu. The Lumbar Traction menu offers the following options:

**Traction Rationale** - Button provides a description of the traction being used as well as definitions of terms used in areas of the Treatment Screen.

**Traction Technique** - Button provides visual steps as to how to set up a patient for lumbar supine or lumbar prone traction.

**NOTE:** After viewing the steps for set up, the Forward Arrow button will automatically go to the Treatment Screen.

TOC






## OPERATING CONTROLS - HOME SCREEN (CONTINUED)

**Edit** - Button provides access to the Treatment Screen to set up parameters for treatment.

**Back and Home** - Buttons go back to the Home Screen.

**Rope Release** - Releases the Traction Cord from the unit.

### sEMG Traction (Model 4798 Only)



The sEMG Traction button goes to a menu to select the desired body area. Press the desired body area button to go to the desired body area + sEMG Traction menu.

The desired body area + sEMG Traction menu offers the following options:

**Traction Rationale** - Button provides a description of the traction being used as well as definitions of terms used in areas of the Treatment Screen.

**Traction Technique** - Button provides visual steps as to how to set up a patient for the desired body area + sEMG traction.

**NOTE:** After viewing the steps for set up, the Forward Arrow button will automatically go to the Treatment Screen.


**NOTE:** There are no traction techniques for the Hip Body Area.

**Edit** - Button provides access to the Treatment Screen to set up parameters for treatment and to the sEMG Graph to set up the sEMG target.

**Back and Home** - Buttons go back to the Home Screen.

**Rope Release** - Releases the Traction Cord from the unit.

### Clinical Protocols™



The Clinical Protocols™ button goes to a menu to select the desired body area. Press the desired body area button to go to the desired body area menu.

**NOTE: FOR CERVICAL AND LUMBAR ONLY** - There are numerous options to choose from before entering the Cervical or Lumbar Traction menu: **Cervical Only**- Whiplash, Headache; **Lumbar Only**- Spondylolisthesis, Strain/Spasm, Sacro Iliac Strain; **Both Cervical and Lumbar**- Radiculopathy Mild, Radiculopathy Moderate, Radiculopathy Severe, Myelopathy, Spondylosis and Facet Joint Syndrome.

The desired body area menu offers the following options:

**Traction Rationale** - Button provides a description of the traction being used as well as definitions of terms used in areas of the Treatment Screen.





# OPERATION

## OPERATING CONTROLS - HOME SCREEN (CONTINUED)

**Traction Technique** - Button provides visual steps as to how to set up a patient for traction treatment of the desired body area.

**NOTE:** After viewing the steps for set up, the Forward Arrow button will automatically go to the Treatment Screen.

**NOTE:** There are no traction techniques for the Hip Body Area.

**Edit** - Button provides access to the Treatment Screen to set up parameters for treatment.

**Back and Home** - Buttons go back to the Home Screen.

**Rope Release** - Releases the Traction Cord from the unit.

### Multimedia Card (MMC) Graphics Library

The MMC Graphics Library button provides access to visual anatomical libraries of the four body areas: Cervical, Lumbar, Hip and Hand/Wrist.

### Patient Card

The Patient Card button allows access to retrieve previously saved patient treatment information from the Patient Card.


### Rope Release

Releases the Traction Cord from the unit.

**TOC**



## OPERATING CONTROLS - TREATMENT SCREEN



The Touch Screen User Interface allows the operator to access and set up for therapy in the following areas of the Treatment Screen:

**NOTE:** Should you make a mistake while entering data, you may correct by re-pressing the appropriate key and re-entering the data.

**NOTE:** On the Touch Screen User Interface, any parameters surrounded in blue boxes are active and can be changed at any time. Any parameters surrounded in gray boxes are inactive and cannot be changed.

### (1) Traction Meter

Displays the amount of tension being delivered to a patient (in kilograms, pounds, or newtons).

### (2) Time Remaining

Displays the approximate number of minutes that remain in the current traction therapy session.

### (3) Time Remaining During Hold/Rest Cycles

Displays an amount of time when a Hold or Rest cycle is occurring during treatment as well as show the number of minutes that remain in that cycle.

### (4) Treatment Status

Displays one of the following to show the current status of the treatment being performed: Setup, Running, Paused, and Completed.

TOC





# OPERATION

## OPERATING CONTROLS - TREATMENT SCREEN (CONTINUED)

### (5) Home

The Home button will return to the Home Screen of the unit.

### (6) Back

The Back button will return to the previous screen.

### (7) Progressive

There are two buttons to use under the Progressive section. The first button allows you to choose from Static or Intermittent Traction Mode to use during the Progressive phase of the traction therapy session. The second button allows you to enter the step number (1-9) desired in the Progressive phase of the traction therapy session.

**NOTE:** The third box in the Progressive section is an inactive button that displays the approximate amount of time (in minutes) selected for the progressive phase of the traction program.

### (8) TX (Traction)

There are two buttons to use under the TX section. The first button allows you to choose from Static, Intermittent or Cyclic Traction Mode to use during the traction phase of the traction therapy session. The second button allows you to enter the amount of time in minutes (1-99) desired for the traction phase of the traction therapy session.

**NOTE:** The TX Time can be changed during treatment. Press the TX Time button and use the Up or Down Arrows to change the time in one minute increments. Press the Return Arrow to accept changes.

### (9) Regressive

There are two buttons to use under the Regressive section. The first button allows you to choose from Static or Intermittent Traction Mode to use during the Regressive phase of the traction therapy session. The second button allows you to enter the step number (1-9) desired in the Regressive phase of the traction therapy session.

**NOTE:** The third box in the Regressive section is an inactive button that displays the approximate amount of time (in minutes) selected for the regressive phase of the traction program.

### (10) Rope Release

The Rope Release button is used to release the Traction Cord in order to pull out of the unit freely for setup.

To release the cord, either pull on the Traction Cord and it will slowly feed out, or hold the Rope Release button down and pull the Traction Cord out.

**NOTE:** If there is more than 2kg (5 lbs or 19N) of tension, the Rope Release button is disabled.



## CAUTION

- In the event of a loss of power to the unit or when quick release is needed, traction tension should only be released by having the patient move towards the traction head to release the tension on the rope. Once the tension on the rope has been released, loosen the patient harness adjustment straps.

**TOC**





## OPERATING CONTROLS - TREATMENT SCREEN (CONTINUED)

### (11) Speed

The percentage of speed of when the Traction Cord is being pulled or released. There are three speed options to choose from: 30%, 50% or 100%.



### (12) Pre-tension

Pre-tension allows the clinician to set a certain level in weight and time to allow the opportunity to make adjustments to harnesses before treatment starts.



### (13) Edit Hold/Rest Times

Displays the amount of hold and rest times (0-99 seconds) of Progressive, TX and Regressive traction entered during setup. To enter the hold or rest time, press the appropriate button, then enter the desired time, using the number keypad. Press the Return Arrow button to accept.

**NOTE:** The Hold and Rest Time can be changed during the traction therapy session by pressing the

Edit Hold/Rest Times button, press the appropriate button, and use the Up or Down Arrow buttons any changes. Press the Return Arrow button to accept.



### (14) Edit Min/Max Levels

Displays the Minimum and Maximum Traction Tension entered during setup. To enter the Minimum or Maximum Traction Tension, press the Edit Min/Max Levels button, press the appropriate button, then enter the desired setting, using the number keypad. Press the Return Arrow button to accept.

**NOTE:** The Minimum and Maximum Traction Tension can be changed during the traction therapy session by pressing the Edit Min/Max Levels button, press the appropriate button, then use the Up or Down Arrow buttons to make changes in 1 kg (1 lb or 1 N) increments. Press the Return Arrow button to accept.

**TOC**



# OPERATION

## OPERATING CONTROLS - TREATMENT SCREEN (CONTINUED)

### FOR LUMBAR TRACTION ONLY:

The Minimum and Maximum Traction Tension levels can be entered in two different ways: Force Levels in Weight or Percentage of Body Weight.

For Force Levels in Weight, simply enter the Minimum and Maximum levels as described previously.



For Percentage of Body Weight, Press the Body Weight button and enter the body weight of the patient. Press the Max. Level % of BW button to enter the percentage of the patient's body weight to be the maximum traction tension level. Press the Min. Level % of BW button to enter the percentage of the patient's body weight to be the minimum traction tension level.

### (15) Clinical Resources

This button offers the following features of the traction unit: [\(See Patient Data Card and User Protocols in the Operation Section for more information\)](#)

#### • Patient Card

- Patient Card (Patient Name)
- Edit Current Pain Profile
- Edit/Save Completed Pain Profile
- Erase Patient Card

#### • Protocols

- Retrieve Protocol
- Save Protocol
- Clinical Protocols™

#### • Utilities

- Unit Settings [\(See Pages 22 - 23 for more information\)](#)

### (16) Stop

Stops the treatment program. Traction tension slowly decreases to zero.

### (17) Pause

Pauses the treatment program. Traction tension slowly decreases to zero. Press the Pause Button again to resume treatment program and traction tension.

### (18) Start

Starts the treatment program.



## sEMG TRACTION TREATMENT

**NOTE:** sEMG is only used as a trigger to begin traction therapy.


### Preparing for sEMG Traction Therapy



Plug the sEMG lead wire into the front of the unit.

## WARNING

- Do not apply electrodes over broken or compromised skin (e.g., sunburn, cuts, acne) due to increased risk of skin reactions.



Examine the patient's skin for any wounds and clean the skin. DJO, LLC recommends using only Dura-Stick® II Electrodes to obtain the most accurate sEMG feedback. Connect the Dura-Stick II disposable electrodes to the lead wire. Leave the electrodes on the protective backing until treatment area has been prepared. Ensure the electrodes are applied securely to the skin.

## DANGER

- Handle, clean, and dispose of components and accessories that have come in contact with bodily fluids according to National, Local and Facility rules, regulations and procedures.



Give the Patient Interrupt Switch to the patient, instructing that pressing the red button will stop the treatment.


The Patient Interrupt Switch also serves as a visual indicator for the patient with sEMG. Once a patient is hooked up to electrodes, the blue lights on the Patient Interrupt Switch show the amount of tension of the patient. The increased number of blue lights indicate higher levels of tension in a patient.

## WARNING

- The Patient Interrupt Switch must be in the patient's grasp throughout the traction therapy.




## SEMG TRACTION TREATMENT (CONTINUED)



Press the sEMG Traction button from the Home Screen.


**NOTE:** Once the sEMG Traction button is pressed, a continuous audible tone will sound from the Patient Interrupt Switch to indicate sEMG Traction activity readings.



Set the parameters for sEMG Traction treatment, using the following buttons on the Treatment Screen:

### (1) View Graph

The View Graph button allows the clinician to toggle between the Treatment Graph to the sEMG graph.



### (2) sEMG Graph


The sEMG Graph allows the setting target level of muscle activity for triggering traction therapy to start. There are three buttons and a scale to set the sEMG target:



**Target button** - This toggle button allows the clinician to select the target either manually or take the average of muscle activity achieved over a 15 second period.



**Adjust/Capture Target button** - Once the target is selected, this button allows the clinician to either adjust the manual target or capture the average target.



**To Adjust the Manual Target** - Press the Target button until Manual is selected, press the Adjust Target button, press the Up or Down Arrow buttons to adjust the Target to the prescribed level, and press the Return Arrow to accept. The target range is from 5 to 2,000  $\mu$ V.

TOC





## sEMG TRACTION TREATMENT (CONTINUED)



### To Capture the Average Target -

Press the Target button until Average is selected. Press the Capture Target button. Press the Begin Capture button to initiate the 15 second period. Once the 15 second period is completed, the average target of muscle activity is captured, and the Up and Down Arrow buttons can be used to increase (up to 200%) or decrease (down to 5%) the percentage of the captured target value. The Reset button can be pressed to clear the captured target.

**NOTE:** The capture may be stopped at any time during the 15 second period by pressing the End Capture button.



**Alarm button** - This button allows the clinician to choose if the alarm should sound above or below the target.

**Above:** The alarm will sound when the sEMG activity exceeds the target.

**Below:** The alarm will sound when the sEMG activity is below the target.

Press the Start button on the front of the unit to begin treatment.

**NOTE:** Once the Start button is pressed, the continuous audible tone from the Patient Interrupt Switch will stop during treatment.

**NOTE:** When traction treatment begins, if sEMG readings are above target, a message will display saying, **"Ask the patient to relax"**. Once the patient has relaxed and the sEMG readings go below the target, the message will disappear and treatment will start.

Check the sEMG Graph from the Treatment Screen to assure that the sEMG signal is being received.

**NOTE:** The continuous audible tone from the Patient Interrupt Switch will begin again once treatment is complete, if the Patient Interrupt Switch has been pressed during treatment, or if the Stop button is pressed.

Examine the patient's skin again after treatment.





# OPERATION

## STARTING, PAUSING AND STOPPING TREATMENT

**NOTE:** Read and follow steps of the Patient Preparation section on **Pages 24 - 25**, the Operating Controls of the Home and Treatment Screens on **Pages 26 - 32**, and the sEMG Traction section on **Pages 33 - 35**, as well as learn Operating Controls of the Treatment Screen to set up traction treatment.

### When the Start button is pressed:

- the audible tone sounds
- A message will display saying, **"Beginning traction. Ensure table is unlocked. Touch anywhere to continue"**.

**NOTE:** If the screen is not touched, the message will disappear in 3 seconds.

- the Traction Cord tightens
- traction therapy begins
- the Treatment Screen displays the maximum traction tension, the minimum traction tension, the hold time, the rest time, the current tension, the time remaining in the traction therapy, and the running status
- the Treatment Graph of the traction therapy becomes shaded as it goes through the treatment

**NOTE:** Monitor the traction therapy closely.

### When the Pause button is pressed:

- the audible tone sounds
- traction therapy is paused
- the Traction Cord will release
- the tension slowly decreases to zero

**NOTE:** The Pause button may be used to adjust the patient's harness or position during treatment.

Resume the traction therapy by pressing the Pause button again.

Traction therapy can be stopped at any time by pressing the Stop button or by pressing the red button of the Patient Interrupt Switch.

### When the Stop button is pressed:

- traction tension will decrease gradually
- the Traction Cord will release
- the audible tone sounds
- the Treatment Screen will return to Setup Mode

If necessary, check the patient's harness and position on the table, and change the traction parameters. Restart traction therapy by pressing the Start button.

### When the red button of the Patient Interrupt Switch is pressed:

- the audible tone sounds
- the motor is stopped
- a warning will display saying *The Patient Switch has been pressed. The treatment has been terminated.* Touch the screen to acknowledge the warning and the tension will decrease to zero
- the Traction Cord will release

### If the unit loses power:

- the motor stops



## CAUTION

- In the event of a loss of power to the unit or when quick release is needed, traction tension should only be released by having the patient move towards the traction head to release the tension on the rope. Once the tension on the rope has been released, loosen the patient harness adjustment straps.

TOC





# OPERATION

## STARTING, PAUSING AND STOPPING TREATMENT (CONTINUED)

When the traction therapy is finished, the audible tone sounds, the traction tension will decrease gradually, the Traction Cord will release, and a message will display saying, "*Treatment has been completed. Do you want to save data to Patient Card?*" Touch the Yes or No button to respond. **(See Saving Treatment to a Patient Data Card for more information.)**

**NOTE:** The beeper will beep every 30 seconds until the screen is touched to acknowledge the message.

**NOTE:** Wait until all tension has released automatically before attempting to remove the harness from the patient. Release the patient from the traction harness according to the manufacturer's instructions.

**TOC**



# OPERATION

## SAVING TREATMENT TO A PATIENT DATA CARD


The Triton traction unit incorporates a Patient Data Card interface that allows transfer of patient therapy data from the unit to the Patient Data Card. The unit allows storage and recall of the following patient session data onto the Patient Data Card: all therapy session parameters, before and after Patient Pain Profiles, and Session Notes (stored on card via PC only). Each Patient Data Card can store multiple sessions and each session can be recalled within the unit.

### New Patient Data Card Setup




Insert a new Patient Data Card (with the gold chip facing up) into the Patient Data Card port located on the front of the traction unit.

**NOTE:** Insert the Patient Data Card before treatment.




Press the Clinical Resources button on the front of the traction unit. Press the Edit Current Pain Profile button located in the Patient Card section of the menu.

Enter the following information under the Pain Before Treatment column:



**Pain Scale** - There are eleven pain scale settings to choose from a numeric scale of 0 to 10. Press the arrows left to right from what the patient describes pain from being No Pain (0) to Worst Pain Possible (10).

**NOTE:** If the patient is unsure of rating pain from a numeric scale, a visual pain scale is available. Toggle the middle button between the left and right arrows to choose between a Numeric or a Visual Analog Scale. Press the arrows left to right to move the visual bar from what the patient describes pain from being No Pain to Worst Pain Possible.



**Pain Type** - There are eleven pain types to choose from based on how the patient describes pain: Numbing, Dull Ache, Throbbing, Pulsing, Tingling, Nagging, Pinching, Burning, Shooting, Stabbing, Radiating, or None Selected.

TOC






# OPERATION

## SAVING TREATMENT TO A NEW PATIENT DATA CARD (CONTINUED)




**Pain Map** - Press the Edit Pain Map button. Press the area of the body to highlight where the patient describes pain. Each time an area of the body diagram is pressed, a highlighted square is produced. Press the area again to remove the highlighted square.




**NOTE:** The Clear button clears all highlighted areas of the body diagram.

Press the Back button or Home button to begin set up and to perform the patient therapy session.



**NOTE:** All pain information can be edited during a patient therapy session by pressing the Clinical Resources button and Edit Current Pain Profile button. However, it is recommended to enter all pre-treatment pain information **before** the patient therapy session.



Once a therapy session has been completed, press the Clinical Resources button. Press Edit/Save Completed Pain Profile button to enter all post-treatment pain information **after** the therapy session.

**(See previous instructions for Pain Scale, Pain Type, and Pain Map and enter information under the Pain After column)** to show any progress made from the treatment.

**NOTE:** Once pain information is entered, review all information to ensure of accuracy. **Once saved, pain information cannot be changed.**

Press the Save to Patient Card button to save patient therapy session. This will save all session parameters, and pain information.



# OPERATION

## SAVING TREATMENT TO A NEW PATIENT DATA CARD (CONTINUED)

After the Save to Patient Card button is pressed, a keyboard will be displayed to enter the patient's name, if the card is blank. Once the patient's name is entered, press the Return Arrow to accept.

Approximately ten treatments can be saved on a Patient Data Card. Use one card per patient.

A message will be displayed saying *treatment has been saved to the Patient Card for (patient's name)*. Touch the screen to acknowledge. This will return you to the Treatment Screen.

Remove the Patient Data Card for filing with patient records.






# OPERATION

## RETRIEVING DATA FROM EXISTING PATIENT DATA CARD




### Existing Patient Data Card Use

Insert the Patient Data Card (with the gold chip facing up) into Patient Data Card port located on the front of the traction unit.




Press the Clinical Resources button on the front of the traction unit. Press the Patient Card (Patient's Name) button located in the Patient Card section of the menu.

**NOTE:** If the Patient Data Card is not inserted, this button is deactivated.



The treatment list is located on the right side. Press the Up or Down Arrow button to select the desired treatment. Press the Return Arrow to accept. This will show the following information: Before and After Treatment Pain Information (Pain Map, Pain Scale and Pain Type) and Session Notes.



**NOTE:** The box located on the left side of the treatment list is a review of treatment parameters of the selected treatment. The review of the treatment shows the following information: Patient's Name, Date, Start Treatment Time, End Treatment Time, Treatment Parameters and Session Notes (stored on card via PC only).

Press Run This Treatment button.

Read and follow the steps of the Patient Preparation section on [Pages 24 - 25](#) to set up traction treatment.

Press the Start button to begin treatment.







## ERASING PATIENT DATA CARD



Insert the Patient Data Card (with the gold chip facing up), to be erased, into the Patient Data Card port located on the front of the traction unit.




When the Patient Data Card is erased, a message will be displayed saying, **"Patient Card has been erased"**. Touch the screen to acknowledge the message and it will return to the Clinical Resources Screen.



Press the Clinical Resources button on the front of the traction unit.

Press the Erase Patient Card button located in the Patient Card section of the menu.



A message will display asking, **"Are you sure you want to erase the Patient Card?"** Press Yes or No.




# OPERATION

## USER PROTOCOLS

This library is a series of protocols created by the user and stored in the unit memory. The following information gives general instructions as to setup, saving and access of User Protocols. Should the Restore Default Protocol button be pressed, through the Utilities section of the Clinical Resources Screen, all User Protocols will be permanently removed from the unit.

**NOTE:** Twenty User Protocols can be saved per body part as well as a Default Protocol for Cervical and Lumbar only.



### Setup and Saving User Protocols




Select and edit the parameters of the desired treatment from the Treatment Screen.

Press the Clinical Resources button on the front of the traction unit. Press the Save Protocol button located in the Protocols section of the menu. Press the Up or Down Arrow button to select the number of User Protocol to save the desired treatment to or the Default Protocol button to save treatment as the default. Press the Return Arrow to accept.

**NOTE:** The Default Protocol is only available for Cervical or Lumbar Traction.



After the Return Arrow is pressed, a keyboard will be displayed to enter the name of the User Protocol. Once the User Protocol name has been entered, press the Return Arrow to accept.




A message will be displayed saying **Current treatment settings have been saved as "Protocol Name"**. Touch the screen to continue.

# OPERATION


## USER PROTOCOLS (CONTINUED)

### Accessing User Protocols




Press the Clinical Resources button on the front of the traction unit.

Press the Retrieve Protocol button located in the Protocols section of the menu.



Select the body part for the desired User Protocol to retrieve.



The list of the User Protocols is located on the right side and the body part buttons are located on the left side to toggle between body part User Protocols. Press the Up or Down button to review the User Protocols.


Select the desired User Protocol by pressing the Return Arrow.




# OPERATION

## CLINICAL PROTOCOLS™


The Clinical Protocols™ section is a series of traction protocol presets where the body area is selected by the user and the Clinical Protocols™ algorithm will select the parameter settings. These Clinical Protocols™ are to be used only as guidelines. Each patient should be individually assessed by a licensed practitioner to determine the appropriateness of the protocol parameters prior to use. All Clinical Protocols™ can be edited to suit appropriate patient treatment prescription and patient comfort. The following information gives general instructions to access, select and setup Clinical Protocols™. Each Clinical Protocol can be set up and edited in the same basic manner.



Press the Clinical Protocols™ button from the Home Screen.



Select the desired body area for a Clinical Protocol.




There are numerous options to choose from for **Cervical and Lumbar only**:

- |                      |                      |
|----------------------|----------------------|
| <b>Cervical Only</b> | <b>Lumbar Only</b>   |
| • Whiplash           | • Spondylolisthesis  |
| • Headache           | • Strain/Spasm       |
|                      | • Sacro Iliac Strain |

**Both Cervical and Lumbar**

- Radiculopathy Mild
- Radiculopathy Moderate
- Radiculopathy Severe
- Myelopathy
- Spondylosis
- Facet Joint Syndrome



The next screen will show the body area menu of the Clinical Protocol.


TOC






# OPERATION


## CLINICAL PROTOCOLS™ (CONTINUED)



Press the Traction Rationale button to view the text explaining the rationale for the type of traction associated with the Clinical Protocol selected as well as a definition of terms associated to the Treatment Screen.



Press the Traction Technique button to view patient preparation instructions.  
**NOTE:** There are no traction techniques for the Hip body area.



Press the Edit button to enter the Maximum and Minimum Traction Levels or to change any of the Clinical Protocol presets.









# OPERATION

## MULTIMEDIA CARD (MMC) GRAPHICS LIBRARY

The Multimedia Card (MMC) Graphics Library offers an anatomical library that is designed to aid the operator in visually identifying and recognizing specific skeletal structures and muscles.



Press the MMC Graphics Library button from the Home Screen.



Select the button of the desired body area.

Once the body area is selected, a list of related items to the body area will be displayed. Choose from the following:

### Lumbar

- Bones
- Muscles
- Scoliosis
- Sacroiliac Joint
- Degenerative Disc
- Radicular Pain
- Herniated Disc
- Sciatica
- Spinal Stenosis
- Ankylosing Spondylitis

### Cervical

- Bones
  - The Spine
  - The Cervical Spine
- Herniation
- Tension Headache
- Anterior Cervical Muscles
- Whiplash
- Whiplash II

### Hand/Wrist

- Bones
- Ligaments
- Palmer View
- Muscles
- Tendons
- Tendonitis
- Carpal Tunnel
  - CTS: Cross Section
  - CTS
  - CTS: Palmer View
- Dupuytren's Contracture Scarring

### Hip

- Bones
- Muscles
- Hip: Bursitis





# OPERATION

## DTS PULL PATTERNS (IF APPLICABLE)

The DTS Pull Patterns, only available with P/N 2842 Triton DTS Card, offer a selection of various intermittent traction programs. It is up to the clinician to determine the applicability and appropriate parameters of the presets as applied to the patient.

### Installing DTS Software Upgrade Card



Turn the traction unit off with the On/Off Switch on the back of the unit.

Remove the Multimedia Card (MMC) from the front of the traction unit.

**NOTE:** The Multimedia (MMC) Card can be discarded once removed.


Insert the Triton DTS Card into the Multimedia Card (MMC) Port.




Turn the traction unit on with the On/Off Switch on the back of the unit.

A message will display on the Touch Screen User Interface saying, **"Unit has been changed to a DTS unit. Touch anywhere to continue."**


### Selecting a DTS Pull Pattern



From the Home Screen, press the DTS Pull Pattern button.



Choose the desired body area for the DTS Pull Pattern.



Select one of the five desired DTS Pull Patterns.

**NOTE:** The DTS Pull Pattern parameters can be edited and saved as the default Pull Pattern or as a User Protocol. (See the User Protocol Section on Pages 43 - 44 for more information)

TOC





## ACCESSORIES

The following provides the users of the Triton traction unit the necessary information to order the replacement accessories most commonly used with the system. This list of replacement accessories are designed for use with Chattanooga traction devices. When ordering, provide the respective part number, description and quantity desired.

Ref.	Description
7040	Saunders Cervical Traction Device
2825	Quickwrap DTS Belting System
48039	Traction Clevis
48083	Patient Interrupt Switch
48084	Patient Interrupt Switch with sEMG
48018	Patient Switch Hook Kit
48031	Replacement Traction Cord
27465	Patient Data Cards (25/pack)
2768	Patient Data Management System
27321	sEMG Lead Wire
4766	sEMG Module
2842	Triton DTS Card
48129	Multimedia (MMC) Card
42192	32 mm Round Dura-Stick II Electrodes

Ref.	Description
42061	Dura-Stick II Electrodes - 3 cm (1.25 in) Round (40/case)
1301	TX Cervical Pillow without Cover
1341	TX White Cotton Pillow with Cover
10889	TX Pillow Cover



## ACCESSORIES

Ref.	Description
21284	Euro Mains Power Cord
78121	US Equivalent Mains Power Cord
20971	Australian Mains Power Cord
20972	Swiss Mains Power Cord
20973	UK Mains Power Cord
20974	Danish Mains Power Cord
20975	Japanese Mains Power Cord
20976	Indian Mains Power Cord
20977	Israeli Mains Power Cord

**NOTE:** The Mains Power Cord shipped with the System will accommodate the electrical requirements for the country of use  
**Not to be sold in the US.**





# TROUBLESHOOTING

The Triton traction unit is designed with patient safety in mind. An error can be caused by both external and internal disturbances. Errors can be caused by disruptions in the power supply (such as a voltage break), and excess or inadequate voltage. Errors may also be caused by patient movement during the traction therapy.

All traction unit errors and warnings are categorized by three numerical digits: Messages beginning with a 1 are general errors, a 2 symbolizes internal warnings and errors and a 3 signifies a critical error where the problem has locked up the unit. **(See Warning on Page 54 for more information)**

Before calling for service, carefully review this User Manual. In the event you are still unable to correct the problem, contact your DJO, LLC dealer for all repair service. Be certain to specify your model number, serial number, and a detailed description of the issue you encountered.

Problem	Possible Remedy
<b>The Power On/Off Switch is in the “ON” position - nothing happens.</b>	<ul style="list-style-type: none"> <li>• Verify the Mains Power Cord is properly connected to the power supply outlet.</li> <li>• Contact dealer or DJO, LLC for service.</li> </ul>
<b>The LED Indicator (Power On/Off) Light glows, but there is nothing displayed on the LCD, or the LCD displays confusing wording.</b>	<ul style="list-style-type: none"> <li>• Verify the voltage from the power supply of the electrical outlet is the same as that is listed on your Voltage Rating and Serial Number Plate.</li> </ul>
<b>The LED Indicator (Power On/Off) blinks - nothing is displayed on the screen.</b>	<ul style="list-style-type: none"> <li>• The unit is in Screen Saver Mode. Touch the screen or any button to reactivate the unit.</li> </ul>
<b>The Patient Interrupt Switch is not working.</b>	<ul style="list-style-type: none"> <li>• Make sure the Patient Interrupt Switch is properly connected to the Patient Interrupt Switch Receptacle.</li> </ul>
<b>The Rope Release will not work.</b>	<ul style="list-style-type: none"> <li>• Release all tension on the Traction Cord.</li> <li>• If there is more than 2kg (5 lbs or 19N) of tension, the Rope Release button is disabled.</li> <li>• If patient is attached, have patient move towards the traction head to release tension.</li> <li>• Shake the Traction Cord while holding down the Rope Release button.</li> <li>• Turn unit off and return to DJO, LLC for repair.</li> </ul>





# TROUBLESHOOTING

Problem	Possible Remedy
<b>Unable to properly read Patient Data Card.</b>	<ul style="list-style-type: none"> <li>• Properly insert a Patient Data Card.</li> <li>• Use a known good Patient Data Card.</li> <li>• If problem persists, contact dealer or DJO, LLC for service.</li> </ul>
<b>Attempted to use an Invalid Patient Data Card.</b>	<ul style="list-style-type: none"> <li>• Properly insert a Patient Data Card.</li> <li>• Use a known good Patient Data Card.</li> <li>• If problem persists, contact dealer or DJO, LLC for service.</li> </ul>
<b>No Session Data is available on the inserted Patient Data Card.</b>	<ul style="list-style-type: none"> <li>• Save session data to Patient Data Card.</li> <li>• Use a known good Patient Data Card.</li> <li>• If problem persists, contact dealer or DJO, LLC for service.</li> </ul>
<b>Patient Data Card is full.</b>	<ul style="list-style-type: none"> <li>• Purchase additional Patient Data Cards from an authorized DJO, LLC dealer.</li> </ul>
<b>sEMG Traction button is grayed out.</b>	<ul style="list-style-type: none"> <li>• An sEMG Module is required to initiate EMG Traction</li> <li>• Make sure sEMG Module is properly inserted into the Triton traction unit. If problem persists, contact dealer or DJO, LLC for service.</li> </ul>





# TROUBLESHOOTING

## ERROR MESSAGES

ERROR CODE	ERROR TYPE	DEFINITION	PROBABLE CAUSES	POSSIBLE REMEDY
100	<b>WARNING</b>	Treatment has been running for 8 seconds, but no tension is detected on the rope.	Too much slack in Traction Cord.	Remove Traction Cord slack.
101	<b>WARNING</b>	Patient pressed Patient Interrupt Switch.	Patient Interrupt Switch button pressed.	Touch the Touch Screen to clear the message.
102	<b>WARNING</b>	Patient Interrupt Switch is unplugged.	Patient Interrupt Switch not properly connected to unit.	Properly connect Patient Interrupt Switch and touch the Touch Screen to clear message.
104	<b>WARNING</b>	User selected Patient Card button on Utilities Screen, but no traction treatments were found on the card.	Wrong or bad Patient Data Card inserted into unit.	Insert correct Patient Data Card.
105	<b>WARNING</b>	User selected Save To Card, but no card is inserted into the unit.	No Patient Data Card is inserted.	Properly insert correct Patient Data Card into unit.
106	<b>WARNING</b>	User selected Save To Card, but the card currently inserted is not a Patient Card.	Wrong or bad Patient Data Card inserted into unit.	Insert correct Patient Data Card.
107	<b>WARNING</b>	User selected Save To Card, but the card currently inserted is full.	Memory full on Patient Data Card used.	Save Data to PDMS and erase Patient Data Card.
108	<b>WARNING</b>	User typed in a blank Patient Name.	No Patient ID assigned.	Enter Patient ID.
112	<b>WARNING</b>	Error upgrading Control Board Software.	Unknown	Contact Dealer or DJO, LLC for Service.
113	<b>WARNING</b>	Error upgrading Motor Board Software.	Unknown	Contact Dealer or DJO, LLC for Service.
114	<b>WARNING</b>	User pressed START, but has not yet set the Max Level value to a value greater than zero.	User has not completed Treatment Setup.	Complete Treatment Setup prior to pressing Start.
115	<b>WARNING</b>	User pressed HOME button from treatment review screen while a treatment was running.	Home button has been pressed.	Press Stop To End Treatment prior to pressing Home.





# TROUBLESHOOTING

## ERROR MESSAGES (CONTINUED)



### WARNING

- In the event that an Error message or Warning appears beginning with a 2 or 3, immediately stop all use of the system and contact the dealer or DJO, LLC for service. Errors and Warnings in these categories indicate an internal problem with the unit that must be tested by DJO, LLC or a Field Service Technician certified by DJO, LLC before any further operation or use of the unit. Use of a unit that indicates an Error or Warning in these categories may pose a risk of injury to the patient, user or extensive internal damage to the unit.





# MAINTENANCE

## CLEANING

**NOTE:** Before cleaning, disconnect the unit from the power source. Periodically, clean the system with a clean, lint free cloth moistened with water and mild antibacterial soap. If a more sterile cleaning is needed, use a cloth moistened with an antimicrobial cleaner.

Do not submerge the system in water. Should the unit accidentally become submersed, contact the dealer or DJO, LLC Service Department immediately. Do not allow cleaning solutions or water to enter the ventilation holes in the unit. This could permanently damage the unit. Do not attempt to use a unit that has been wet inside until inspected and tested by a Field Service Technicians certified by DJO, LLC.

### Touch Screen Cleaning

Clean unit display lens using a soft damp cloth, moistened with warm water and soap if necessary. DO NOT USE alcohol or chlorine based solvents as this may damage the display.

## PREVENTIVE MAINTENANCE SCHEDULE

The design of this device is mechanical in nature and thereby has moving parts that may become worn or require lubrication from time to time. It is recommended that this device be placed on a periodic maintenance schedule to examine for lubrication needs or replacement of components. Please reference the Service Manual for particular parts and maintenance. The schedule of the maintenance depends upon the frequency and duration of device usage and should be determined by the user.

## CALIBRATION REQUIREMENTS

Annual factory calibration is required for Triton traction units. The unit should be sent to the factory or a Field Service Technician certified by DJO, LLC for this procedure.

The Triton Traction unit and/or the Triton DTS Traction unit purchased by your facility (models 4749 or 4798), is calibrated during the manufacturing process. The device is ready to be placed into service upon delivery.





# MAINTENANCE

## SERVICE

Should the Triton Traction Unit require service, contact the selling dealer or DJO, LLC Service Department.

All returned units to the factory for service must include the following:

### WARRANTY REPAIR/OUT OF WARRANTY REPAIR

1. Written statement containing the following information;

- RA Number - Obtain from Factory
- Unit Model Number
- Unit Serial Number
- Contact person with Phone and Fax Numbers
- Billing Address (for Out of Warranty Repair)
- Shipping Address (Where to Ship Unit after Repair)
- Detailed Description of Problem or Symptoms

2. Copy of original invoice issued at purchase of the unit.

3. Ship unit to Factory in the original container with all accessories and information as required in item one (1) above to:

DJO, LLC  
Chattanooga Repair Center  
47492 SD Hwy 22  
PO Box 709  
Clear Lake, SD 57226 USA

When shipping the unit to the dealer or factory, make certain the original packaging is used. If the original packaging is not available, contact DJO, LLC to obtain the packaging materials listed on [Page 17](#) for shipment. Any damage sustained from improper packaging may render the warranty null and void.

Service to these units should be performed only by a Field Service Technician certified by DJO, LLC.

The Triton Traction unit and the Triton DTS Traction unit Service Manuals are available for purchase and can be requested from the selling dealer or DJO, LLC Service Department. The Service Manual contains safety precautions, nomenclature, specifications, troubleshooting, removal and replacement instructions, general maintenance, calibration instructions, parts lists, schematics, warranty and other information which would assist a certified service technician to repair the unit.





# WARRANTY

DJO, LLC (“Company”) warrants that the Triton Traction Unit (“Product”) is free of defects in material and workmanship. This warranty shall remain in effect for two years (24 months) from the date of original consumer purchase. If this Product fails to function during the two year warranty period due to a defect in material or workmanship, at the Company’s option, the Company or the selling dealer will repair or replace this Product without charge within a period of thirty (30) days from the date on which the Product is returned to the Company or the dealer.

All repairs to the Product must be performed by a service center authorized by the Company. Any modifications or repairs performed by unauthorized centers or groups will void this warranty.

The warranty period for the accessories shipped with the unit is 180 days. Accessories consist of the materials shipped with the unit.

**This Warranty Does Not Cover:**

Replacement parts or labor furnished by anyone other than the Company, the selling dealer or a certified Company service agent.

Defects or damage caused by labor furnished by someone other than Company, the selling dealer or a certified Company service agent.

Any malfunction or failure in the Product caused by product misuse, including, but not limited to, the failure to provide reasonable and necessary maintenance or any use that is inconsistent with the Product User Manual.

**COMPANY SHALL NOT BE LIABLE IN ANY EVENT FOR INCIDENTAL OR CONSEQUENTIAL DAMAGES.**

Some locations do not allow the exclusion or limitation of incidental or consequential damages, so the above limitation or exclusion may not apply to you.

To obtain service from Company or the selling dealer under this warranty:

1. A written claim must be made within the warranty period to the Company or the selling dealer. Written claims made to the Company should be sent to:

DJO, LLC  
1430 Decision St  
Vista, CA 92081 USA  
T: 1-800-592-7329 USA  
T: 1-317-406-2209  
F: 1-317-406-2014  
chattgroup.com

2. The Product must be returned to the Company or the selling dealer by the owner.

This warranty gives you specific legal rights and you may also have other rights which vary from location to location.

The Company does not authorize any person or representative to create for it any other obligation or liability in connection with the sale of the Product. Any representation or agreement not contained in the warranty shall be void and of no effect.

**THE FOREGOING WARRANTY IS IN LIEU OF ALL OTHER WARRANTIES, EXPRESSED OR IMPLIED, INCLUDING ANY WARRANTY OR MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE.**





# Moving Rehabilitation Forward™



ISO 13485 Certified

DJO, LLC  
1430 Decision St  
Vista, CA 92081 USA  
T: 1-800-592-7329 USA  
T: 1-317-406-2209  
F: 1-317-406-2014  
chattgroup.com

© 2011 DJO, LLC

48144\_F



TOC

